



The LASER ONE is a portable natural gas leak analyser with LASER spectroscopy. This highly sensitive detector offers all the advantages of optical detection in a compact and portable device. The HUBERG LASER ONE gives you precise and reliable findings when detecting methane leaks.

2355 Bishop Circle West Dexter, MI 48130, USA 800.624.2026 734.995.2547









## LASER ONE



POWER SUPPLY			
Battery type	Li-ion rechargeable battery pack (user replaceable)	Battery charger	Separate intelligent battery charger powered from mains supply (100-240V, or via vehicle power outlet)
Battery life	Typical 10 hours from fully charged (with backlight activated)	Charge time	Approximately 4.5 hours from complete discharge
GAS RANGES			
Gases Measured	CH <sub>4</sub>	Range	1-9,999ppm / 1%-100% (Vol)
Response time	Without Probe : T <sup>90</sup> = 2 seconds		
	With Probe : T <sup>90</sup> = 3.5 seconds		
Accuracy	<b>0.1ppm-10.9ppmVol:</b> +/-0.7ppmVol	11-9,999ppmVol: +/- 7% of reading	<b>1%-100%Vol</b> +/- 5% of reading
Resolution	0.1ppm / 0.1% (Range dependent)	Minimum detectable limit	0.3ppm
PUMP			
Flow	Minimum: 0.6 L/Min	Typical: 0.8 L/Min	Maximum: 1.0 L/Min
ENVIRONMENTAL CONDITIONS	S		
Operating Temperature	-25 to 50°C / -13 to 122° F	Humidity Range	5% to 80% Relative Humidity (non -condensing)
Atmospheric Pressure/ Range	1013mbar +/-100mbar/29.91 inHg +/- 2.95inHg	Protection Level	IP65
PHYSICAL			
Weight	1.3 kg / 2.9 lb	Enviromental Storage Conditions	Humidity <95% Temperature -40°C to +60°C/ -40°F to 140°F
Dimensions	229 x 97 x 109 mm / 9 x 4 x 4.2 inches		
Alarm Sound Level	~65 dB (A) @30cm / 12"	User Interface	• Display (LxH): 86 x 47mm / 3.39" x 1.85"
			Jogdail: Scroll menu for rapid and easy selection of menu items.
			• 3 Function Keys
Gas Connections	Quick Release Coupler	Electrical Connections	Multiway adaptor for battery charger and USB communication port with a PC
		Carry Strap	30mm/1.2" wide synthetic woven strap
CERTIFICATION RATING			



©2021 Product designs and specifications are subject to change without notice. User is responsible for determining suitability of product.













